





### DECLARATION OF CONFORMITY

## MCD Medical Line THA.leia essential

| 2000076M | MCD Medical Line THA.leia essential |
|----------|-------------------------------------|
|          |                                     |

## MCD Medical Computers Deutschland GmbH Konrad-Zuse-Ring 17 A/B 41179 Moenchengladbach

- Germany -

hereby declare in sole responsibility that the above mentioned system(s) comply with the essential requirements of the following EC Directives as well as EN60601-1:2006 + Cor.:2010+A1:2013, if used for its intended purpose:

For Models w/o (1) EMC-Directive 2014/30/EU

For Models with

WiFi-Module:

(3) RED-Directive 2014/53/EU

#### Applied harmonized standards or normative documents:

DIN EN 60601-1-2:2016-05, EN 61000-4-2-2009-12, EN 61000-4-3:2006 + A1:2008 +

(1) & (2) A2:2010, EN 61000-4-4-2013-04, EN 61000-4-5-2019-03, EN 61000-4-6-2014-08, EN

61000-4-8:2010

EN60950-1:2006+A11:2009+A12:2011+AC:2011+A2:2013,

(3) EN50566:2013, EN301489-1 v1.9.2, EN301489-17 v2.2.1,

EN300328 v1.9.1, EN301893 v1.8.1, EN300440-2 v1.4.1

#### Power supply:

Magic Power EN 62366:2008+A1, EN 60601-1-6:2010+A1, EN 60601-1:2006+A11+A1+A12,

ATM090T-P120 IEC 62366:2007+A1, IEC 60601-1-6:2010+A1, IEC 60601-1:2005+A1,

Certificate No. TA 50372234

This declaration certifies compliance with the guidelines mentioned, but does not include any guarantee of properties. The safety instructions in the supplied product documentation must be observed.









DECLARATION OF CONFORMITY



# **EU RoHS / REACH Statement**

MCD feels responsible for the environment and the health of user, patients and others as well. Therefore MCD products are designed an manufactured in order to comply with currently valid RoHS directive (2011/65/EC) / (EU 2015/863) and the REACH regulation (1907/2006) mentioned below.

#### RoHS:

We hereby declare in sole responsibility that the mentioned above products are in full compliance with EU Directive 2011/65/EC and EU 2015/863 with respect to the following substances:

| Substance                      | Element | Limit value (%) |
|--------------------------------|---------|-----------------|
| Lead                           | Pb      | 0,1             |
| Mercury                        | Hg      | 0,1             |
| Cadmium                        | Cd      | 0,01            |
| Hexavalent chromium            | Cr (VI) | 0,1             |
| Polybrominated biphenyls       | PBB     | 0,1             |
| Polybrominated diphenyl ethers | PBDE    | 0,1             |
| Bis(2-ethylhexyl) phthalate    | DEHP    | 0,1             |
| Butyl benzyl phthalate         | BBP     | 0,1             |
| Dibutyl phthalate              | DBP     | 0,1             |
| Diisobutyl phthalate           | DIBP    | 0,1             |

#### **REACH:**

MCD Medical Computers Deutschland GmbH is a manufacturer of pc systems for medical environments.

We fulfill the REACH Regulation (Regulation (EC) no. 1907/2006) related to the requirements with relevance and continuously monitor changes of the REACH Regulation and the list of candidates.

We will inform you in the context of our business relationship about changes of delivered products caused by the EC REACH Regulation and will, in the individual case, coordinate suitable measures with you. With reference to article 33 of the REACH ordinance we inform you about the following:

An up-to-date list of candidates (updated June 10th, 2022, last check for topicality November 15th 2022) in accordance with article 59 (1, 10) of the ordinance (EC) no. 1907/2006 (REACH) was published.

The mentioned above products and their packaging do not contain any substances of this current list of candidates in concentrations of more than 0.1 mass %.

Moenchengladbach, .2022-11-15

(Place and date of the exhibit)

<u>Thomas Hollex</u>

(Issuer)

TÜVRheinland

Management System EN ISO 13485:2016 www.tuv.com ID 9105055459